



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2021FA0006

Date: 15Sep2021

Urgent Field Safety Notice
Inferior Vena Cava (IVC) Filter

For Attention of: Chief Executive / Risk Management / Purchasing

Contact details of local representative (name, e-mail, telephone, address etc.)

Cook Medical Europe Ltd.
O'Halloran Road
National Technology Park
Limerick, Ireland
E-mail: European.FieldAction@CookMedical.com
Phone: Please refer to the attached Country Contacts List

For any further information or support concerning the information within this FSN please contact your local Cook Medical Sales Representative or Cook Medical Europe Ltd.

FSN & FSCA Ref: 2021FA0006



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2021FA0006

Urgent Field Safety Notice (FSN)

Inferior Vena Cava (IVC) Filter

Risk addressed by FSN

1. Information on Affected Devices	
1.	<p>1. Device Type(s)</p> <p>The CE-marked Günther Tulip Vena Cava Filter Set and the Cook Celect Platinum Vena Cava Filter Set are in the scope of this FSN:</p> <p>The Günther Tulip Vena Cava Filter Set (IGTCFS-65-2-UNI-TULIP) includes the Günther Tulip Vena Cava Filter implant and the introducer system components. The Günther Tulip IVC filter implant is composed of a paramagnetic cobalt chromium alloy (50 mm long when compressed to a diameter of 30 mm) and is supplied preloaded on a femoral filter introducer. A jugular introducer, introducer system, and pre-dilator are also supplied. The Günther Tulip Vena Cava Filter implant is designed to act as a permanent or retrievable filter. The Günther Tulip Vena Cava Filter implant may be retrieved if clinically indicated; the IFU provides more information about optional filter retrieval.</p> <p>The Cook Celect Platinum Vena Cava Filter Sets (IGTCFS-65-2-FEM/JUG/UNI(-FT)-CELECT-PT) includes the Cook Celect Platinum Vena Cava Filter implant and the introducer system components. The Cook Celect Platinum Vena Cava filter implant is composed of a paramagnetic cobalt chromium alloy (49 mm long when compressed to a diameter of 30 mm) with platinum markers and is supplied preloaded on a femoral or jugular filter introducer. An introducer system, and pre-dilator are also supplied. The Cook Celect Platinum Vena Cava Filter implant is designed to act as a permanent or retrievable filter. The Cook Celect Platinum Vena Cava Filter implant may be retrieved if clinically indicated; the IFU provides more information about optional filter retrieval.</p>
1.	<p>2. Commercial name(s)</p> <p>Günther Tulip® Vena Cava Filter Set for Femoral and Jugular Vein Approach, Cook Celect® Platinum Vena Cava Filter Set for Femoral Vein Approach, Cook Celect® Platinum Vena Cava Filter Set for Jugular Vein Approach, Cook Celect® Platinum Vena Cava Filter Set for Femoral and Jugular Vein Approach Cook Celect® Platinum NavAlign Uniset Vena Cava Filter Set</p>

FSN & FSCA Ref: 2021FA0006



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2021FA0006

1.	3. Primary clinical purpose of device(s)																				
<p>The Günther Tulip and Cook Celect Platinum Vena Cava Filters are intended to capture blood clots traveling in the infrarenal inferior vena cava in the clinical situations detailed in the Indications for Use section of the IFU.</p> <p>The Günther Tulip and Cook Celect Platinum Vena Cava Filter implants may be retrieved if clinically indicated, see the "Optional Filter Retrieval" section of the IFU for more information.</p>																					
1.	4. Device Model/Catalogue/part number(s)																				
<table border="1"> <thead> <tr> <th data-bbox="268 779 715 817">Product code - RPN</th> <th data-bbox="722 779 842 817">GPN</th> <th data-bbox="850 779 1484 817">Description</th> </tr> </thead> <tbody> <tr> <td data-bbox="268 828 715 929">IGTCFS-65-2-UNI-TULIP</td> <td data-bbox="722 828 842 929">G52926</td> <td data-bbox="850 828 1484 929">Günther Tulip® Vena Cava Filter Set for Femoral and Jugular Vein Approach</td> </tr> <tr> <td data-bbox="268 940 715 1041">IGTCFS-65-2-FEM-CELECT-PT</td> <td data-bbox="722 940 842 1041">G34501</td> <td data-bbox="850 940 1484 1041">Cook Celect® Platinum Vena Cava Filter Set for Femoral Vein Approach</td> </tr> <tr> <td data-bbox="268 1052 715 1153">IGTCFS-65-2-JUG-CELECT-PT</td> <td data-bbox="722 1052 842 1153">G34310</td> <td data-bbox="850 1052 1484 1153">Cook Celect® Platinum Vena Cava Filter Set for Jugular Vein Approach</td> </tr> <tr> <td data-bbox="268 1164 715 1265">IGTCFS-65-2-UNI-CELECT-PT</td> <td data-bbox="722 1164 842 1265">G34504</td> <td data-bbox="850 1164 1484 1265">Cook Celect® Platinum Vena Cava Filter Set for Femoral and Jugular Vein Approach</td> </tr> <tr> <td data-bbox="268 1276 715 1377">IGTCFS-65-2-UNI-FT-CELECT-PT</td> <td data-bbox="722 1276 842 1377">G35581</td> <td data-bbox="850 1276 1484 1377">Cook Celect® Platinum NavAlign Uniset Vena Cava Filter Set</td> </tr> </tbody> </table>				Product code - RPN	GPN	Description	IGTCFS-65-2-UNI-TULIP	G52926	Günther Tulip® Vena Cava Filter Set for Femoral and Jugular Vein Approach	IGTCFS-65-2-FEM-CELECT-PT	G34501	Cook Celect® Platinum Vena Cava Filter Set for Femoral Vein Approach	IGTCFS-65-2-JUG-CELECT-PT	G34310	Cook Celect® Platinum Vena Cava Filter Set for Jugular Vein Approach	IGTCFS-65-2-UNI-CELECT-PT	G34504	Cook Celect® Platinum Vena Cava Filter Set for Femoral and Jugular Vein Approach	IGTCFS-65-2-UNI-FT-CELECT-PT	G35581	Cook Celect® Platinum NavAlign Uniset Vena Cava Filter Set
Product code - RPN	GPN	Description																			
IGTCFS-65-2-UNI-TULIP	G52926	Günther Tulip® Vena Cava Filter Set for Femoral and Jugular Vein Approach																			
IGTCFS-65-2-FEM-CELECT-PT	G34501	Cook Celect® Platinum Vena Cava Filter Set for Femoral Vein Approach																			
IGTCFS-65-2-JUG-CELECT-PT	G34310	Cook Celect® Platinum Vena Cava Filter Set for Jugular Vein Approach																			
IGTCFS-65-2-UNI-CELECT-PT	G34504	Cook Celect® Platinum Vena Cava Filter Set for Femoral and Jugular Vein Approach																			
IGTCFS-65-2-UNI-FT-CELECT-PT	G35581	Cook Celect® Platinum NavAlign Uniset Vena Cava Filter Set																			
2 Reason for Field Safety Corrective Action (FSCA)																					
2.	1. Description of the product problem																				
<p>The purpose of this Field Safety Notice (FSN) is to inform you about updated product labeling (specifically, updated Instructions for Use) for the William Cook Europe ApS Günther Tulip Vena Cava Filter Set and Celect Platinum Vena Cava Filter Set.</p> <p>The IFU updates are described in the table below. The updates are not related to device safety, device performance, or product design changes. The updated information is not reflective of newly identified hazards and/or harms or of a change in risk profile of the devices. Rather, the added information reflects well-known safety information associated with endovascular procedures requiring anesthesia and contrast media.</p>																					

FSN & FSCA Ref: 2021FA0006



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2021FA0006

Summary of Labeling Updates:	
Section of IFU	Description of Changes
Device Description	Further clarify that the product is intended for percutaneous placement via a femoral or jugular vein in adult patients
Intended Use / Indications for Use	Modified to better reflect existing clinical evidence.
Contraindications	<ol style="list-style-type: none"> Updated to contraindicate use in minors/pediatric patients and pregnant patients. Note: Use in these patient populations was previously addressed in a Precaution statement; therefore, this update reinforces previously communicated information. The IFU for the Günther Tulip IVC filter implant now includes a Contraindication for use in vena cava below 15 mm in diameter, aligning with existing Celect Platinum use specifications.
Warnings and Precautions	Clarified language and added new warnings and precautions to provide further emphasis related to existing topics in the IFU.
Potential Adverse Events	Aligned with the available post-market surveillance evidence. No new potential adverse events were added. One potential adverse event (coagulopathy) was removed from the list.
How Supplied	Text was added to mitigate the risk of resterilization of the final device.
2.	<p>2. Hazard giving rise to the FSCA</p> <p>No specific feedback regarding device use, device safety or device performance gave rise to this update. Rather, the updates to device labeling were to ensure alignment with ongoing regulatory requirements and best practices.</p> <p>The target population for the Günther Tulip and Celect Platinum Vena Cava Filter Sets remains unchanged; specifically, these devices are intended for patients at risk for PE. However, the IFU updates includes contraindications for two specific patient groups (i.e., minors/pediatrics and pregnant women). While healthcare professionals may assess the potential benefit of IVC filter placement to outweigh the potential risk in these patients, this update reinforces the fact that safety and performance of the Günther Tulip and Celect Platinum IVC filter implants have not been established in these patients</p>

FSN & FSCA Ref: 2021FA0006



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2021FA0006

3. Type of Action to mitigate the risk			
3.	<p>1. Action To Be Taken by the User</p> <p><input checked="" type="checkbox"/> Take note of amendment/reinforcement of Instructions for Use (IFU)</p> <p><input checked="" type="checkbox"/> Other</p> <ol style="list-style-type: none"> No retrospective action for previously implanted products is warranted The electronic versions of the IFUs can be found on the Cook Medical Web https://ifu.cookmedical.com/ifuPub/searchIfu.jsf by Catalogue Number (RPN) search A Cook Medical Sales Representative will personally follow-up and provide corrected IFUs for customer's inventory. Please complete the Customer Response Form within 5 business days of receiving the Field Safety Notice and return it to Cook Medical as directed on the form. 		
3.	<p>2. Particular considerations for: Implantable device</p> <p>Is follow-up of patients or review of patients' previous results recommended? No Compliance with current routine follow-up guidance is recommended.</p>		
3.	<table border="1"> <tr> <td>3. Is customer Reply Required? (If yes, form attached specifying deadline for return)</td> <td style="text-align: center;">Yes</td> </tr> </table>	3. Is customer Reply Required? (If yes, form attached specifying deadline for return)	Yes
3. Is customer Reply Required? (If yes, form attached specifying deadline for return)	Yes		
3.	<p>4. Action Being Taken by the Manufacturer</p> <p><input checked="" type="checkbox"/> IFU or labelling change <input checked="" type="checkbox"/> Other</p> <p>Customers will be contacted by a Cook Medical Sales Rep for the purpose of swapping IFUs from old IFUs to new updated IFUs on all impacted unused devices in the customers possession.</p>		


4. General Information					
4.	<table border="1"> <tr> <td>1. FSN Type</td> <td>New</td> </tr> <tr> <td>2. Further advice or information already expected in follow-up FSN?</td> <td>No</td> </tr> </table>	1. FSN Type	New	2. Further advice or information already expected in follow-up FSN?	No
1. FSN Type	New				
2. Further advice or information already expected in follow-up FSN?	No				
4.	<p>3. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)</p> <table border="1"> <tr> <td>a. Company Name</td> <td>William Cook Europe</td> </tr> <tr> <td>b. Address</td> <td>Sandet 6 4632 Bjaeverskov Denmark</td> </tr> </table>	a. Company Name	William Cook Europe	b. Address	Sandet 6 4632 Bjaeverskov Denmark
a. Company Name	William Cook Europe				
b. Address	Sandet 6 4632 Bjaeverskov Denmark				

FSN & FSCA Ref: 2021FA0006



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

FSN & FSCA Ref: 2021FA0006

4.	4. The Competent (Regulatory) Authority of your country has been informed about this communication to customers.	
4.	5. List of attachments/appendices:	Reply form Country Contacts List
4.	6. Name/Signature	 Lissi Walmann Manager, Regulatory Reporting, Quality Assurance William Cook Europe

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.</p>

FSN & FSCA Ref: 2021FA0006



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

Field Action Customer Reply Form

1. Field Safety Notice (FSN) information	
FSN Reference number	2021FA0006
FSN Date	15Sep2021
Product/ Device name	Günther Tulip® Vena Cava Filter Set and Cook Celect® Platinum Vena Cava Filter Set
Product Code(s)	1. IGTCFS-65-2-UNI-TULIP 2. IGTCFS-65-2-FEM-CELECT-PT 3. IGTCFS-65-2-JUG-CELECT-PT 4. IGTCFS-65-2-UNI-CELECT-PT 5. IGTCFS-65-2-UNI-FT-CELECT-PT
Batch/Serial Number (s)	Please refer to the attached list of affected lot numbers

2. Customer Details	
Account Number	
Healthcare Organisation Name	
Organisation Address	
Contact Name	
Telephone number	
Email	

3. Customer action undertaken on behalf of Healthcare Organisation	
<input type="checkbox"/>	I confirm receipt of the Field Safety Notice and that I read and understood its content.
<input type="checkbox"/>	The information and required actions have been brought to the attention of all relevant users and executed.
<input type="checkbox"/>	I have affected unused devices for IFU swap - enter Lot number and quantities in table below.
Print Name	
Signature	
Date	

4. Return acknowledgement to sender	
Email	European.FieldAction@CookMedical.com
Customer Helpline	Please refer to the attached Country Contacts List
Fax	+ 353 61 239294
Deadline for returning the customer reply form	Please return this form within 5 business days of receipt



COOK MEDICAL EUROPE LTD.
O'HALLORAN ROAD
NATIONAL TECHNOLOGY PARK
LIMERICK, V94 N8X2, IRELAND
TEL: +353 61 334440 FAX: +353 61 334441
WWW.COOKMEDICAL.EU

If you have any unused affected products, please indicate the part number, lot number and quantity:

Product Part Number	Product Lot Number	Quantity

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective actions.